



INTERNATIONAL FIRST CERTIFICATION

SERTİFİKA

Bu sertifika,

ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SANAYİ TİCARET ANONİM ŞİRKETİ

HAS EMEK SANAYİ SİTESİ 1469. CADDE NO:10 İVEDİK OSB YENİMAHALLE ANKARA

kuruluşunun,

MASKE İMALAT, İTHALAT VE İHRACAT

EA 4 kapsamında,

ISO 9001:2015

Kalite Yönetim Sistemi standardının şartlarına uyan bir yönetim sistemi kurduğunu ve uyguladığını onaylamak üzere verilmiştir.

İlk Yayın Tarihi	: 21.09.2020
Yayın Tarihi	: 21.09.2020
Sertifika Geçerlilik Tarihi	: 3 Yıl / 20.09.2023
Sertifika Bitiş Tarihi	: 20.09.2021
Sertifika No	: IFC-Q - 09 - 20 - 1592



Onay
21.09.2018/Rev 01



Kalite Yönetim Sistemi
TSEN ISO/IEC 17021-1
AB-0138-YS



IFC GLOBAL SERTİFİKASYON MUAYENE VE EĞİTİM HİZMETLERİ ANONİM ŞİRKETİ

Adalet Mah. Manas Blv. No:39 / 2203 - Folkart Towers / Bayraklı İZMİR TÜRKİYE T: +90 850 304 35 00 F: +90 850 304 35 00

Bu belge, müşterinin IFC'nin kurallarına ve sözleşme şartlarına uyduğu süreç geçerlidir. Sertifika geçerlilik durumu IFC internet sitesinden takip edilebilir.

www.ifcglobal.com.tr info@ifcglobal.com.tr



INTERNATIONAL FIRST CERTIFICATION

SERTİFİKA

Bu sertifika,

ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SANAYİ TİCARET ANONİM ŞİRKETİ

HAS EMEK SANAYİ SİTESİ 1469. CADDE NO:10 İVEDİK OSB YENİMAHALLE ANKARA

kuruluşunun,

MASKE İMALAT, İTHALAT VE İHRACAT

EA 4 kapsamında,

ISO 14001:2015

Çevre Yönetim Sistemi standardının şartlarına uyan bir yönetim sistemi kurduğunu ve uyguladığını onaylamak üzere verilmiştir.

İlk Yayın Tarihi : 21.09.2020
Yayın Tarihi : 21.09.2020
Sertifika Geçerlilik Tarihi : 3 Yıl / 20.09.2023
Sertifika Bitiş Tarihi : 20.09.2021
Sertifika No : IFC-E - 09 - 20 - 1592



Onay

FR 030/01/2018/Rev.01

IFC GLOBAL SERTİFİKASYON MUAYENE VE EĞİTİM HİZMETLERİ ANONİM ŞİRKETİ

Adalet Mah. Manas Blv. No:39 / 2203 - Folkart Towers / Bayraklı İZMİR TÜRKİYE T: +90 850 304 35 00 F: +90 850 304 35 00

Bu belge, müşterinin IFC'nin kurallarına ve sözleşme şartlarına uyduğu sürece geçerlidir. Sertifika geçerlilik durumu IFC internet sitesinden takip edilebilir.

www.ifcglobal.com.tr info@ifcglobal.com.tr



INTERNATIONAL FIRST CERTIFICATION

SERTİFİKA

Bu sertifika,

ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SANAYİ TİCARET ANONİM ŞİRKETİ

HAS EMEK SANAYİ SİTESİ 1469. CADDE NO:10 İVEDİK OSB YENİMAHALLE ANKARA

kuruluşunun,

MASKE İMALAT, İTHALAT VE İHRACAT

kapsamında,

OHSAS 18001:2007

İş Sağlığı ve Güvenliği Yönetim Sistemi standardının şartlarına uyan bir yönetim sistemi kurduğunu ve uyguladığını onaylamak üzere verilmiştir.

İlk Yayın Tarihi	: 21.09.2020
Yayın Tarihi	: 21.09.2020
Sertifika Geçerlilik Tarihi	: 3 Yıl / 20.09.2023
Sertifika Bitiş Tarihi	: 20.09.2021
Sertifika No	: IFC- O 09 - 20 - 1592

Onay

FR 240/01-01.2018/Rev.01

IFC GLOBAL SERTİFİKASYON MUAYENE VE EĞİTİM HİZMETLERİ ANONİM ŞİRKETİ

Adalet Mah. Manas Blv. No:39 / 2203 - Folkart Towers / Bayraklı İZMİR TÜRKİYE **T: +90 850 304 35 00** **F: +90 850 304 35 00**

Bu belge, müşterinin IFC'nin kurallarına ve sözleşme şartlarına uyduğu sürece geçerlidir. Sertifika geçerlilik durumu IFC internet sitesinden takip edilebilir.

www.ifcglobal.com.tr info@ifcglobal.com.tr

CERTIFICATE of Registration



*This is to Certify that the
Medical Devices -- Quality Management System*

of

ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SANAYİ TİCARET ANONİM ŞİRKETİ

**İVEDİK ORGANİZE SANAYİ BÖLGESİ HAS EMEK SANAYİ SİTESİ 1469.
SOKAK NO:10 YENİMAHALLE /ANKARA / TÜRKİYE**

**has been independently assessed and is compliant
with the requirements of**

ISO 13485:2016

This Certificate is applicable to the following product or service ranges:

PRODUCTION OF STERILE AND NON STERILE DISPOSABLE SURGERY MASK

STERİL VE NON STERİL TEK KULLANIMLIK CERRAHİ MASKE ÜRETİMİ

Certificate No.: TR52672H

Date of initial registration 17 June 2020

Date of this Certificate 17 June 2020

Surveillance audit on or before 16 June 2021

Recertification Due / Certificate expiry 16 June 2023

This Certificate is property of Staunchly Management & System Services Ltd. and remains valid
subject to satisfactory surveillance audits.

Director

STAUNCHLY MANAGEMENT & SYSTEM SERVICES LTD.

Suite 48, 88-90 Halton Garden, London, EC1N 8PN.

Phone : +44 345 680 0199

Email : info@staunchlyservices.com Web : www.staunchlyservices.com

SMS/F109A/17/REV02

For precise and updated information concerning the present certificate mail to info@staunchlyservices.com

This Certificate is the property of Staunchly Management & System Services Private Limited and shall be returned immediately when demanded



TEST REPORT
DENEY RAPORU

AB-0583-T

20033391-
ing

09-20

Customer name: ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SAN. TİC.
Address: Has Emek Sanayi Sitesi 1469. Cadde No: 10 İvedik OSB Yenimahalle/
ANKARA
Buyer name: -
Contact Person: FÜSUN ANLADI CEYLAN
Order No: -
Article No: MARKA: ALVENT MODEL: MA-250
Name and identity of test item: Blue non-woven medical mask.
The date of receipt of test item: 11.09.2020
Re-submitted/re-confirmation date: -
Date of test: 11.09.2020-21.09.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 5

The Turkish Accreditation Agency (TURKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports.

EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.

The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.



Date
21.09.2020

Customer Representative
Tuğba AKTAŞ

Head of Testing Laboratory
Sevim A. RAZAK
21.09.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**


AB-0583-T

20033391-
ing

09-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TESTS		
Bacterial Filtration Efficiency-BFE	P	Type IIR
Microbial Cleanliness(Bioburden)	P	
PHYSICAL PROPERTIES		
Breathability(Differential Pressure)	P	
Splash Resistance	P	
P: Pass F: Fail R: Refer to retailer technologist. Tests results were evaluated according to EN 14683:2019+AC :2019 Tablo 1 limit values.		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified.If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033391-
ing

09-20

TEST RESULTS

BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: (Bacterial Filtration Efficiency Testing –BFE /Ref: EN 14683:2019+AC:2019 Medical Face Masks, Requirements and Test Methods

A specimen of the mask material is clamped between a impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/min
Test Flow Time	2 minute
Sample Sizes	5 pieces mask
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml)	5×10^5 cfu/ ml
incubation conditions	24 hour, $35^\circ\text{C} \pm 2^\circ\text{C}$
Positive control sample average of number of Bacteria (C)	2.5×10^3 cfu/ ml

RESULTS			
Number of Test Sample	Test Sample (T) Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency (% B)	Requirement BFE (%)
1	45	%98.2	Type I ≥ 95 Type II ≥ 98
2	36	%98.6	
3	40	%98.4	
4	50	%98.0	
5	33	%98.7	

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033391-
ing

09-20

TEST RESULTS

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: EN ISO 11737-1:2018

The sample is put in extraciton liquid after shaking well, inoculated on the agar.

After incubation at 30 ± 1 ° C for 72 hours, growth microorganisms are counted on the agar.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	16 cfu/g	≤ 30 cfu/g Type I and Type II mask

*cfu= Colony forming unit.

SPLASH RESİSTANCE (ONLY FOR TYPE IIR)

Test Metod: EN 14683:2019+AC :2019 (Clause 5.2.4) the resistance of the medical face mask to penetration of splashes of liquid shall conform to the minimum value given for Type IIR in Table 1

ISO 22609 :2004 Clothing for protection against infectious agents — Medical face masks — Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

6 different samples were taken

	<u>SPLASH RESISTANCE PRESSURE (kPa)</u>	<u>RESULTS</u>	<u>REQUIREMENT</u>
1	>21.3 kPa	PASS	≥ 16 kPa
2	>21.3 kPa	PASS	
3	>21.3 kPa	PASS	
4	>21.3 kPa	PASS	
5	>21.3 kPa	PASS	
6	>21.3 kPa	PASS	
Average Result	>21.3 kPa	PASS	

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033391-
ing

09-20

TEST SONUÇLARI

BREATHABILITY (Differential Pressure)

Test Method: EN 14683:2019+AC :2019 (TS EN 14683+AC:2019) Annex-C

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

Test area is 25 mm in diameter , 5 different sample was taken


Adjusted airflow is 8 l/min. The differential pressure is read directly using a differential pressure manometer .

SAMPLE	DIFFERENTIAL PRESSURE RESULT	REQUIREMENT
1	25.1 Pa/cm ²	< 60 Pa/cm ²
2	29.8 Pa/cm ²	
3	28.9 Pa/cm ²	
4	29.2 Pa/cm ²	
5	27.1 Pa/cm ²	
Average Result	28.0 Pa/cm ²	



EU declaration of conformity

According to 93/42/EEC



Declaration of conformity no - Revision : **DoC002-00**

Technical File No - Rev: **ALF202013-00**

Legal Manufacturer	Altinkaya Elektronik Cihaz Kutuları San. Tic. A.S. 1469. cadde No:10 İvedik OSB 06378 Yenimahalle / ANKARA Turkey Tel: +90 312 963 1613 mask@alvent.com.tr
Product Brand	ALVENT
Product Name	TYPE IIR Earloop Surgical Mask with meltblown layer
Product Range	Disposable Surgical Face Mask
Product Code	MA-250
Medical Device Classification	Class 1 non-sterile rule 1
GMDN class	35177 surgical face mask, single use
UNSPSC Code	42131713 Surgical Mask

We herewith declare under our sole responsibility that the above mentioned product bear the CE Marking, and meet the provisions of the following Council Directive(s) as transposed into national laws.

General Applicable Directives	Council Directive 93/42/EEC of 14 June 1993 (MDD 93/42/EEC) as amended
Annex 93/42/EEC	VII
Notified Body	N/A
Harmonised Standards	EN 14683:2019+AC-EQV
Applicable standards	EN ISO 11737-1 :2018 EN ISO 10993-1 :2009 EN ISO 9073-4 EN ISO 29073-3

All supporting documentation is retained at the manufacturer's premises

Authorized Signature

Ahmet Altınışik
Managing Director

Ankara
Place

24.09.2020
Date

TEST REPORT
DENEY RAPORU

AB-0583-T

20033390-
ing


09-20

Customer name: ALTINKAYA ELEKTRONİK CİHAZ KUTULARI SAN. TİC.
Address: Has Emek Sanayi Sitesi 1469. Cadde No: 10 İvedik OSB Yenimahalle/
ANKARA
Buyer name: -
Contact Person: FÜSUN ANLADI CEYLAN
Order No: -
Article No: MARKA: ALVENT MODEL: MA-356
Name and identity of test item: Blue non-woven medical mask.
The date of receipt of test item: 11.09.2020
Re-submitted/re-confirmation date: -
Date of test: 11.09.2020-21.09.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 5

The Turkish Accreditation Agency (TÜRKAK) is signatory to the multilateral agreements of the European co-operation for the Accreditation (EA) and of the International Laboratory Accreditation (ILAC) for the Mutual recognition of test reports.

EKOTEKS LABORATUVAR ve GÖZETİM HİZMETLERİ A.Ş. accredited by TÜRKAK under registration number [AB-0583-T] for ISO 17025:2017 as test laboratory.

The test and/or measurement results, the uncertainties (if applicable) with confidence probability and test methods are given on the following pages which are part of this report.



Date
21.09.2020

Customer Representative
Tugba AKTAS

Head of Testing Laboratory
Sevim A. RAZAK

21.09.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**


AB-0583-T

20033390-
ing

09-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TESTS		
Bacterial Filtration Efficiency-BFE	P	Type IIR
Microbial Cleanliness(Bioburden)	P	
PHYSICAL PROPERTIES		
Breathability(Differential Pressure)	P	
Splash Resistance	P	
P: Pass F: Fail R: Refer to retailer technologist. Tests results were evaluated according to EN 14683:2019+AC :2019 Tablo 1 limit values.		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified.If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033390-
ing

09-20

TEST RESULTS

BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: (Bacterial Filtration Efficiency Testing –BFE /Ref: EN 14683:2019+AC:2019 Medical Face Masks, Requirements and Test Methods

A specimen of the mask material is clamped between a impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

Test Flow Rate	28,3 L/min
Test Flow Time	2 minute
Sample Sizes	5 pieces mask
Microorganism	<i>Staphylococcus aureus</i> ATCC 6538
Bacterial concentration (cfu/ ml)	5x10 ⁵ cfu/ ml
incubation conditions	24 hour, 35°C ± 2°C
Positive control sample average of number of Bacteria (C)	2.74x10 ³ cfu/ ml

RESULTS			
Number of Test Sample	Test Sample (T) Number of Bacteria (cfu/ml)	Bacterial Filtration Efficiency (% B)	Requirement BFE (%)
1	48	%98.2	Type I ≥95 Type II ≥98
2	35	%98.7	
3	51	%98.1	
4	43	%98.4	
5	40	%98.5	

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033390-
ing

09-20

TEST RESULTS

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: EN ISO 11737-1:2018

The sample is put in extraciton liquid after shaking well, inoculated on the agar.

After incubation at 30 ± 1 ° C for 72 hours, growth microorganisms are counted on the agar.

	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	12 cfu/g	≤ 30 cfu/g Type I and Type II mask

*cfu= Colony forming unit.

SPLASH RESİSTANCE (ONLY FOR TYPE IIR)

Test Metod: EN 14683:2019+AC :2019 (Clause 5.2.4) the resistance of the medical face mask to penetration of splashes of liquid shall conform to the minimum value given for Type IIR in Table 1

ISO 22609 :2004 Clothing for protection against infectious agents — Medical face masks — Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

6 different samples were taken

	<u>SPLASH RESİSTANCE PRESSURE (kPa)</u>	<u>RESULTS</u>	<u>REQUIREMENT</u>
1	>21.3 kPa	PASS	≥ 16 kPa
2	>21.3 kPa	PASS	
3	>21.3 kPa	PASS	
4	>21.3 kPa	PASS	
5	>21.3 kPa	PASS	
6	>21.3 kPa	PASS	
Average Result	>21.3 kPa	PASS	

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

20033390-
ing

09-20

TEST SONUÇLARI

BREATHABILITY (Differential Pressure)

Test Method: EN 14683:2019+AC :2019 (TS EN 14683+AC:2019) Annex-C

Test Condition (21 ± 5) °C ve (85 ± 5) % relative humidity, 4 hrs

Test area is 25 mm in diameter , 5 different sample was taken

Adjusted airflow is 8 l/min. The differential pressure is read directly using a differential pressure manometer .

SAMPLE	DIFFERENTIAL PRESSURE RESULT	REQUIREMENT
1	19.8 Pa/cm ²	< 60 Pa/cm ²
2	21.5 Pa/cm ²	
3	24.0 Pa/cm ²	
4	21.5 Pa/cm ²	
5	22.6 Pa/cm ²	
Average Result	21.8 Pa/cm ²	



EU declaration of conformity

According to 93/42/EEC



Declaration of conformity no - Revision : **DoC003-00**

Technical File No - Rev: **ALF202014-00**

Legal Manufacturer	Altinkaya Elektronik Cihaz Kutuları San. Tic. A.S. 1469. cadde No:10 İvedik OSB 06378 Yenimahalle / ANKARA Turkey Tel: +90 312 963 1613 mask@alvent.com.tr
Product Brand	ALVENT
Product Name	Surgical Face Mask Type IIR Softloop
Product Range	Disposable Surgical Face Mask
Product Code	MA-356
Medical Device Classification	Class 1 non-sterile rule 1
GMDN class	35177 surgical face mask, single use
UNSPSC Code	42131713 Surgical Mask

We herewith declare under our sole responsibility that the above mentioned product bear the CE Marking, and meet the provisions of the following Council Directive(s) as transposed into national laws.

General Applicable Directives	Council Directive 93/42/EEC of 14 June 1993 (MDD 93/42/EEC) as amended
Annex 93/42/EEC	VII
Notified Body	N/A
Harmonised Standards	EN 14683:2019+AC-EQV
Applicable standards	EN ISO 11737-1 :2018 EN ISO 10993-1 :2009 EN ISO 9073-4 EN ISO 29073-3

All supporting documentation is retained at the manufacturer's premises

Authorized Signature

Ahmet Altınışık
Managing Director

Ankara
Place

24.09.2020
Date